
Percent Error from Expected Intensity as a Function of Position of Calibration Pulse in Gate for Tor101d
% Error from Expected Intensity
16%
14%
12%
10%
8%
6%
4%
2%
0%
-2%
0
10
20
30
40
50
60
70
80
90
100
Time Between Trailing Edge of Pulse and End of Gate (usec)
y = -0.0015x + 0.1498
R2 = 0.9988